

#### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

#### LORENZ CALIBRATION, INC. 10 N 738 Prairie Xing Elgin, IL 60124

Zachary Lorenz Phone: 847 845 7400

#### **CALIBRATION**

Valid To: May 31, 2025 Certificate Number: 3435.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,5</sup>:

#### I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Surface Plate – Flatness <sup>3</sup> Repeat Reading <sup>3</sup>	Up to (72 x 144) in Up to 0.002 in	(31 + 1.6 <i>DL</i> ) μin 33 μin	Autocollimator  Repeat reading gage

<sup>&</sup>lt;sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

(A2LA Cert. No. 3435.01) Revised 04/23/2025

Page 1 of 1

<sup>&</sup>lt;sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

This laboratory performs field calibration activities for this parameter. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>&</sup>lt;sup>4</sup> In the statement of CMC, *DL* is the diagonal length of the unit under test in inches.

<sup>&</sup>lt;sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# **Accredited Laboratory**

A2LA has accredited

## LORENZ CALIBRATION, INC.

Elgin, IL

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the R205 – Specific Requirements:

Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system

(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 13th day of June 2023.

Mr. Trace McInturff, Vice President, Accreditation Services

For the Accreditation Council

Certificate Number 3435.01

Valid to May 31, 2025 Revised April 23, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.